



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/044,969	01/15/2002	Yousuke Moriuchi	026575-065	2991

7590

12/27/2005

Platon N. Mandros
BURNS, DOANE, SWECKER & MATHIS, L.L.P.
P.O. Box 1404
Alexandria, VA 22313-1404

EXAMINER

EREZO, DARWIN P

ART UNIT

PAPER NUMBER

3731

DATE MAILED: 12/27/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

88

Office Action Summary	Application No. 10/044,969	Applicant(s) MORIUCHI ET AL.	
	Examiner Darwin P. Erez	Art Unit 3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 06 October 2005.
 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-4, 6, 7 and 16-30 is/are pending in the application.
 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
 5) ☐ Claim(s) _____ is/are allowed.
 6) ☒ Claim(s) 1-4, 6, 7 and 16-30 is/are rejected.
 7) ☐ Claim(s) _____ is/are objected to.
 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 a) ☐ All b) ☐ Some * c) ☐ None of:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
 * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –


(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

2. Claims 1-3, 6, 16-20, 22, 24-29 are rejected under 35 U.S.C. 102(e) as being anticipated by US 6,113,627 to Jang.

As to claims 1, 17 and 26, Figure 4 of Jang, which is shown below, discloses an expandable stent having a plurality of annular expanding members arranged a predetermined distance apart from each other in an axial direction of the stent and forming a waved element; and a plurality of waved connecting members connecting bottoms of the waved elements; wherein the plurality of annular expanding members are arranged in an axial direction of the stent such that no substantial phase difference existed in the waves of the waved elements (see enclosed Fig.); and each of the waved connecting members has a plurality of waves including a wave formed in the clearance between adjacent annular expanding members and having an amplitude larger than that of the other wave, the largest wave having a height measured circumferentially of the stent (see figure below) the ridge of the waved elements having width measured


Art Unit: 3731

circumferentially of the stent (see figure below), the wave height of the largest wave being larger than the width of the ridge; and wherein a portion of the wave of each of the plurality of waved connecting members is positioned between portions of a single one of the waved elements



As to claims 2, 18, 19, 27 and 28, the ridges depicted above are connected to adjacent ridges. Also, the connecting members can be viewed as connecting the bottom of the waves, as shown below.


Art Unit: 3731



As to claims 3, 20 and 29, the figure above shows the width of each waved connecting member not being larger than $\frac{1}{2}$ of the width of the waved element.

As to claims 6 and 22, the total length of the connecting member is more than 1.3 times the distance between the ridges (see figure above).

As the claim 16, the waved connecting member have at least three waves.



Art Unit: 3731

As to claims 24 and 25, Fig. 4 shows a plurality of annular connecting members, one which could be identified as a third annular expanding member. The annular connecting members have the configuration that is recited in the claims, which is similar to the rejected claim 1 above.

Claim Rejections - 35 USC § 103

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. Claims 4, 7, 21, 23 and 30 are rejected under 35 U.S.C. 103(a) as being unpatentable over Jang.

Jang teaches all the limitations of the claims except for any specific recitation of the width of each of the waves connecting members falling within the range of 0.03-0.08 mm, or the width of the clearance between adjacent annular expanding members falling within the range of 0.4-0.8mm. However, it would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the device of Jang to arrive at said ranges, since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, 220 F.2d 454, 456, 105 USPQ 233, 235 (CCPA 1955).

Response to Arguments

5. Applicant's arguments with respect to claims 1-4, 6, 7, 16-30 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

6. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Darwin P. Erezzo whose telephone number is (571) 272-4695. The examiner can normally be reached on M-F (7:30-4:00).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anhtuan T. Nguyen can be reached on (571) 272-4963.

Art Unit: 3731

The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

de


ANH TUAN T. NGUYEN
SUPERVISORY PATENT EXAMINER
12/22/05